



Gel Content Tester for PEX Products

Description

- Test samples of PEX need to be immersed in boiling solvent and the insoluble part measured. The percentage of this part is degree of crosslinking
- Round glass flask with a heating jacket
- Cooling tower
- Ring stand with suitable clamps and test holders with lids for the sample
- Sample SS304 cage 100 Pcs
- To complete the measurement you will also require a heating oven and analytical balance
- According to ISO 10147 , DIN 16892 , ASTM D 2765
- Chemicals is up to the customer