



Thermal Cycling Test Stand for Pipes and Fittings

Description

- According to EN12293, ISO10508
- Sample cabin
- 3 pressure lines
- Hot and cold water tank
- Chiller
- Computer controlled
- Report in MSWORD
- Pressure up to 14 bar
- Sliding door for sample cabin
- PLC controlled
- Automatic cycle control

AHP PLASTIK MAKINA